

VIEW LAB MONTHLY SUMMARY

| Lab Name | Govt. PHC, Avatty | Details About Robust Analysis | | | | | | | | |
|---|--------------------------------|---|--|--------------------|--------|-------------------------------------|------|-----------------------|-------|------|
| Lab No | 12082 | Detail About Monthly Summary | | | | | | | | |
| Month | March | Detail about SDI | | | | | | | | |
| Year | 2023 | | | | | | | | | |
| Constituent Group | Chemistry III | | | | | | | | | |
| Click on the analyte to view Graphical Data | | | | | | All Analyser Result | | Print | | |
| Date of Result Entered : 20/03/2023 | | | | | | | | | | |
| Date of Report Published : 04/04/2023 | | | | | | | | | | |
| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI | U |
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1878 | 91.48 | 7.46 | 6.82 | 94 mg/dL | 0.37 | 0.31 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1851 | 1.33 | 9.08 | 0.12 | 1.1 mg/dL | -1.90 | 0.01 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1761 | 110.93 | 5.23 | 5.80 | 116 mg/dL | 0.87 | 0.28 |
| SDI Range | | Interpretation | | | | | | | | |
| Within -1.00 to +1.00 | | Excellent. | | | | | | | | |
| Within ±1.01 to ±2.00 | | Good. | | | | | | | | |
| Within ±2.01 to ±2.99 | | Accept with caution. Warning Signal. | | | | | | | | |
| Beyond ±3.0 | | Unacceptable performance. Action Signal. | | | | | | | | |

Links

| | | | |
|--------------------------------|---------------------------------|------------------------------------|--|
| Modify Address | Modify Methods | Submit Lab Results | View Lab Monthly Summary |
| Yearly Summary | Change Password | Home | Sign-out |