

VIEW LAB MONTHLY ALL RESULT SUMMARY

Lab Name **Govt. PHC, Ammankurichy**Lab No **12519**[Back to Search Page](#)Month **September**Year **2023**[Print](#)Constituent Group **Chemistry III**Date of Result Entered : **20/09/2023**Date of Report Published : **05/10/2023**

Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	36	312.41	4.03	12.59			4.20
			Transasia / Erba	18	307.20	6.79	20.85			9.83
			Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	321 mg/dL	0.68	0.69
			Any Analyser (Automation / Semi Automation)	68	312.35	4.82	15.05			3.65
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	38	1.30	7.23	0.09			0.03
			Transasia / Erba	21	1.26	9.52	0.12			0.05
			Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.37 mg/dL	0.76	0.00
			Any Analyser (Automation / Semi Automation)	69	1.31	8.62	0.11			0.03
3	CHOLESTROL III	CHOD-PAP III	Biosystems	35	113.55	3.41	3.87			1.31
			Transasia / Erba	19	113.64	2.93	3.33			1.53
			Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	115 mg/dL	0.32	0.20
			Any Analyser (Automation / Semi Automation)	151	113.39	4.43	5.02			0.82

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

