

VIEW LAB MONTHLY ALL RESULT SUMMARY

Lab Name	Govt. PHC, Ammankurichy									
Lab No	12519	Back to Search Page								
Month	May									
Year	2023	Print								
Constituent Group	Chemistry III									
Date of Result Entered : 19/05/2023										
Date of Report Published : 05/06/2023										
Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	33	179.81	5.48	9.85			3.43
			Transasia / Erba	19	177.53	5.62	9.98			4.58
			Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	187 mg/dL	0.65	0.44
			Any Analyser (Automation / Semi Automation)	69	181.20	5.81	10.52			2.53
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	38	1.78	8.91	0.16			0.05
			Transasia / Erba	22	1.85	7.73	0.14			0.06
			Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.68 mg/dL	-0.61	0.01
			Any Analyser (Automation / Semi Automation)	68	1.75	7.05	0.12			0.03
3	CHOLESTROL III	CHOD-PAP III	Biosystems	36	116.50	6.42	7.47			2.49
			Transasia / Erba	20	115.73	3.06	3.54			1.58
			Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	119 mg/dL	0.76	0.27
			Any Analyser (Automation / Semi Automation)	154	114.84	5.08	5.83			0.94
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

