



**ROOTS INDUSTRIES INDIA LIMITED**  
**ROOTS METROLOGY & TESTING LABORATORY**  
 R.K.G Industrial Estate, Ganapathy, Coimbatore-641 006,India  
 Phone :9585840192/9003432356



**CALIBRATION CERTIFICATE**

ULR Number: CC275219000083790F

Name and address of the customer The Medical officer Government Primary Health Center, Kadiyapatti – Pudhukottai( dist)	<b>Certificate no</b> : Rmtl/01/319107848/18 <b>Date of calibration</b> :03/08/2023 <b>Suggested calibration due</b> : 02/08/2024 <b>Customer po .reference.&amp; date:</b>
--	--

**DESCRIPTION OF DUC**

Name Of The Duc: Pipette Make :Microlux Model :NA Sl.No :NA Range/Size : 10micro liter	Receiving Condition Of Duc :Good Ref. Document :As Per Manufactures Specification Environmental Condition :25.3 Deg & 54.3% Rh Uncertainty Of Measurement :As Per Report Location :lab
--	--

**THE MASTERS ARE TRACEABLE TO NATIONAL/INTERNATIONAL STANDARDS AS PER ISO/IEC 17025.**

**DETAILS OF MASTER USED**

Standard Instrument Used	Make/ Model	Id No	Traceable To/Certificate No	Validity
Multi Product Calibrator	Fluke	Rml-262	Wavecal,Blr/Cs-18/01-01	1-6-2024
Micro balance	mettler toledo/XP26PC	QP-01	B115129039/CC-2523	1-6-2024
Universal Measuring Instrument With Sensor For Environmental Parameter	ALMEMO/2590-4AS	QP-10	H19070544/CC-2052	1-6-2024

**CALIBRATION FIELD: WEIGHT Calibration**

Sl.no	Range	Master readings micro litter	Calibrated values micro litter	Error observed micro litter	Uncertainty
1	10micro litter	10	9.9	0.1	0.055
2	10micro litter	10	9.8	0.1	0.055
3	10micro litter	10	9.7	0.1	0.055
4	10micro litter	10	9.5	0.1	0.055
5	10micro litter	10	9.8	0.1	0.055

remarks

uncertainty estimated for a confidence level of 95% at the k factor=2  
 the reported results are valid only for the condition of the inst./gauge received at the time of and under the  
 condition of the calibration  
 DUC –device under calibration  
 Calibration label /status : Calibration result pass & fail are indicated

 CALIBRATION ENGINEER		 AUTHORIZED SIGN	This Report Refers Only To The Sample Submitted And May Not Be Reproduces In Full Or In Part, Without Written Permission From Roots Metrology & Testing Laboratory, Coimbatore, cell:9585840192
--------------------------	--	---------------------	--