

VIEW LAB MONTHLY ALL RESULT SUMMARY

Lab Name	Govt. PHC, Sathiyamangalam									
Lab No	12504	Back to Search Page								
Month	August									
Year	2023	Print								
Constituent Group	Chemistry III									
Date of Result Entered : 19/08/2023										
Date of Report Published : 01/09/2023										
Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	34	172.03	5.51	9.48			3.25
			Transasia / Erba	23	173.17	4.83	8.36			3.49
			Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	176 mg/dL	0.05	0.42
			Any Analyser (Automation / Semi Automation)	64	174.34	5.46	9.51			2.38
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	35	1.76	6.47	0.11			0.04
			Transasia / Erba	21	1.74	6.34	0.11			0.05
			Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.6 mg/dL	-1.34	0.01
			Any Analyser (Automation / Semi Automation)	62	1.76	7.18	0.13			0.03
3	CHOLESTROL III	CHOD-PAP III	Biosystems	33	114.98	3.90	4.49			1.56
			Transasia / Erba	17	111.42	3.18	3.54			1.72
			Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	107 mg/dL	-1.47	0.22
			Any Analyser (Automation / Semi Automation)	149	113.96	4.36	4.97			0.81
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

