

			CAL	IBRATI	ON	CERT	IFIC	AT	E	
CERTIFICATE NO: SBS/CL/23/13855					MEDICAL DEVICES				CES	age No:1 of 1
Issue Date					14-09-2023					
SRF No & Date					SRF/23/00827-0002 & 13-09-2023					
Receipt Date					13-09-2023					
Calibration Date					13-09-2023					
Calibration Due					12-09-2024					
Cust	omer Name & Addre	55								
	ERNMENT PRIMAR		LTH CENTRE							
	HAPADI-636121,SA			•						
				Details of De	vice L	Inder Calibra	tion (DL	IC)		
Desc	ription :	SEMI A	UTO ANALYZ	FR	Make	& Model			ROBONIK & PRIETEST T	OUCH
Range : MULTI			Sr. No		AT1410211RBK					
Resolution : MULTI			Identification No : NA			NA				
DUC Condition : SATISFACTORY			1	Location				LABORATORY		
				mental Condi			Proce	dure		
nvir	onmental Details		Temperature:	25.6 ° C		Relative Hum	idity		52% RH	
Calibration Procedure No SBS/CP/MD/2			20		Calibration done at			ONSITE		
				Refere	ence !	Standards De	tails			
No Description Make/ S			Make/ SI No:	ake/ SI No:		Certificate No			Validity	
Electrical Safety Analyser R			Rigel Medical & 44L-1059		M-230809-16-4			10-08-202		

ELECTRICAL SAFETY

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)	
1	Insulation Resisitance	Measured values in MΩ	Uncertainty in % (±)	
	>20MΩ	97	13.92	
2	Earth Leakage	Measured values in μA	Uncertainty in % (±)	
-	<5000µAfor B,BF,CF	154	7.7	
2	Enclosure Leakage	Measured values in µA	Uncertainty in % (±)	
3	<500µAfor B.BF,CF	235	7.2	

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer) P.PRASANNA





Chief Executive

Authorised Signatory