

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Siruvachur  
 Lab No 12619  
 Month April  
 Year 2023  
 Constituent Group Chemistry III

[Details About Robust Analysis](#)  
[Detail About Monthly Summary](#)  
[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)  
 Date of Result Entered : 20/04/2023  
 Date of Report Published : 06/05/2023

1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	122 mg/dL	0.35	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2.1 mg/dL	-0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	118 mg/dL	1.10	0.28
Interpretation										
SDI Range										
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

*Handwritten signature*

Links

[https://home.cmcvellore.ac.in/clinac/viewMonthlySummary/viewResult2\\_new.aspx?resDetId=264&noSamples=1&month=4&year=2023&groupId=98&yearGroup=110](https://home.cmcvellore.ac.in/clinac/viewMonthlySummary/viewResult2_new.aspx?resDetId=264&noSamples=1&month=4&year=2023&groupId=98&yearGroup=110)