

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Siruvachur
 Lab No 12619
 Month May
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/05/2023

Date of Report Published : 05/06/2023

[Click on the analyte to view Graphical Data](#)

Sl No	Analyte	Method	Result	Unit	Target	Delta	Delta 2
1	GLUCOSE III	GOD-POD III	1834	mg/dL	180.88	5.19	9.39
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	1854	mg/dL	1.76	7.48	0.13
3	CHOLESTROL III	CHOD-PAP III	1798	mg/dL	114.60	5.08	5.83
SDI Range			Interpretation				
Within -1.00 to +1.00			Excellent.				
Within ±1.01 to ±2.00			Good.				
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.				
Beyond ±3.0			Unacceptable performance. Action Signal.				

Links

https://www.crickalore.ac.in/Clinical/ViewMonthlySummary/viewResult2_new.aspx?resDelt=266&noSamples=1&month=5&year=2023&groupId=9&yearGroup=110