

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Siruvachur
 Lab No: 12619
 Month: July
 Year: 2023
 Constituent Group: Chemistry III

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

Date of Result Entered : 20/07/2023

Date of Report Published : 07/08/2023

[Click on the analyte to view Graphical Data](#)

1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	114 mg/dL	0.77	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	6.1 mg/dL	0.77	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	109 mg/dL	-0.86	0.22
SDI Range Interpretation										
Within -1.00 to +1.00 Excellent.										
Within ±1.01 to ±2.00 Good.										
Within ±2.01 to ±2.99 Accept with caution. Warning Signal.										
Beyond ±3.0 Unacceptable performance. Action Signal.										

Handwritten signature

Links

http://www.cmcvellore.ac.in/diag/ViewMonthlySummary/viewResult2_new.aspx?resDetId=2678&nosamples=1&month=7&year=2023&groupId=9&yearGroup=110