

10/9/23, 2:35 PM

External Quality Assurance Scheme - View Monthly Summary

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Siruvachur  
 Lab No 12619  
 Month August  
 Year 2023  
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/08/2023

Date of Report Published : 01/09/2023

1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	159 mg/dL	-1.84	0.42
2	CREATININE III	Jafes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	111 mg/dL	-0.60	0.22

SDI Range		Interpretation
Within -1.00 to +1.00		Excellent.
Within ±1.01 to ±2.00		Good.
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.
Beyond ±3.0		Unacceptable performance. Action Signal.

*Handwritten signature/initials in green ink.*