



SIRUVACHUR O...









			CAL	IBRATION (CERTI	FICATE	<u> </u>	
CERTIFICATE NO:			SBS/CL/23/13850)	MECHANICAL		P	
ssue Date				14-09-2	14-09-2023			
SRF No & Date				SRF/23	SRF/23/00826-0003 & 13-09-2023			
Receipt Date				13-09-2	13-09-2023			
Calibration Date				13-09-2	13-09-2023			
Calibration Due				12-09-2	12-09-2024			
usto	mer Name 8	Addre	ess		1. 11	-		
SOVE	RNMENT	PRIMA	RY HEALTH CENTR	E,				
SIRU	ACHUR-6	36112,	SALEM DISTRICT.					
				Details of Device Unde	er Calibration	on (DUC)	iii	
Description		:	CENTRIFUGE	Make 8	Model	:	LABOTECH &	
Range		:	3500 RPM	Sr. No	Sr. No		NA	
Resolution		:	1 RPM	Identific	ation No	:	NA	
OUC Condition :		:	Satisfactory	Locatio	Location		LABORATOR	
			Environmenta	l Conditions & Stand	ard Operati	ing Procedure	Details	
Environmental Details Ter				Temperature: 25.6 °C Rel		Relative Humidity		
Calibration Procedure No SBS/				P/ML/04 Calibration		done at ONSITE		
	10			Reference Stan	dards Detai	lls	- t-	
6.No Description				Make/ SI No:	Make/ SI No:		Certificate No	
Digital Tachometer			er	LINE SEIKI / 17	LINE SEIKI / 175-0034V		JRPM-CCTR-A&S-2023-001	

REMARKS

S.No

1

2

3

CALIBRATION RESULTS

DEVICE UNDER

CALIBRATION

RPM

1000

2000

3500

.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.

STANDARD INSTRUMENTS

RPM

999.7


1999.5

3499.4

- .The user should determine the suitability of the instrument for its intended use.
- .The recalibration interval should be determined on the user requirement.
- .The results stated in this certificate relate only to the item calibrated.
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for actor k=2.00.
- Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By rasand (Calibration Engineer) P.PRASANNA





DEVIATION

RPM

0.3

0.5

0.6

Chief Exec
T 11 7 3

EXPANDED UI

SUNSHINE BIOMEDICAL SOLUTIONS