



## CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13851

Page. No : 1 of 1

**Customer Name & Address**

GOVERNMENT PRIMARY HEALTH CENTRE,  
SIRUVACHUR-636112,SALEM DISTRICT.

SRF No.	SRF/23/00826-0005
SRF Date	13-09-2023
Date of Receipt	12-09-2023
Date of Calibration	13-09-2023
Due Date for Calibration	12-09-2024
Issue Date	14-09-2023

**Details of Unit Under Calibration**

Description	Micro Pipette	Make	HANDY PETTE
Range	10-100µl	Model	NA
Resolution	1µl	Material	PVC
Serial Number	NA	Operating Range	10-100µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

**Environmental Condition**

**Calibration Method Used**

Temperature (°C)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

**Standard Used**

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSPL22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

**Result of Calibration in µl**

SI. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	10	9.94	9.97	9.94	9.96	9.93	9.95	-0.05	0.02	0.47
		9.98	9.95	9.92	9.97	9.96				
2	50	49.85	49.86	49.87	49.82	49.83	49.85	-0.15	0.02	0.47
		49.87	49.85	49.82	49.89	49.87				
3	100	99.94	99.97	99.96	99.95	99.98	99.95	-0.05	0.02	0.47
		99.93	99.95	99.97	99.94	99.95				

**Remarks**

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

  
 (Calibration Engineer)  
 MRAGUL



Authorised by:

  
 (Quality Manager/Chief Executive)  
 C.SIVABALAN

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.