

## VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Konapet  
 Lab No 12522  
 Month April  
 Year 2023  
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 19/04/2023

Date of Report Published : 06/05/2023

Sl. No.	Analyte	Method / Principle Name	Analyser Name	No. of Participants	CV	Participants		Total Value
						CV	SD	
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	120 mg/dL
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2.1 mg/dL
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation )	143	111.22	4.31	4.79	114 mg/dL

SDI Range	Interpretation
Within -1.00 to -1.00	Excellent.
Within =1.01 to =2.00	Good.
Within =2.01 to =2.99	Accept with caution. Warning Signal.
Beyond =3.0	Unacceptable performance. Action Signal.

## Links

Search Results	Modify Metrics	Submit Lab Reports	Home
Monthly Summary	Change Password	Home	Home