

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Konapet

Details About Robust Analysis

Lab No 12522

Detail About Monthly Summary

Month August

Detail about SDI

Year 2023

Constituent Group Chemistry III

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered 19/08/2023

Date of Report Published 01/09/2023

Lab No	Analyte	Method	Result	Unit	SDI	CV	CV	CV	CV	CV
12522	GLUCOSE III	GOD-POD III	1841	175.58	5.13	9.00	162	mg/dL	-1.51	0.42
	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	1852	1.76	6.77	0.12	1.9	mg/dL	1.18	0.01
	CHOLESTROL III	CHOD-PAP III	149	113.96	4.36	4.97	114	mg/dL	0.01	0.81

SDI Range	Interpretation
lin -1.00 to -1.00	Excellent
lin -1.01 to -2.00	Good
lin +2.01 to +2.99	Accept with caution Warning Signal
ond +3.0	Unacceptable performance Action Signal

12522	19/08/2023	19/08/2023	19/08/2023