

12522

Detail About Monthly Summary

September

Detail about SDI

2023

Chemistry III

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/09/2023

Date of Report Published : 05/10/2023

Method / Principle Name	Analyser Model	Sample Count	Mean	SD	CV	Target	Delta	Delta %	Delta Z
E III GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	310 mg/dL	-0.06	0.69	
LINE III Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.2 mg/dL	-0.86	0.00	
TROL III CHOD-PAP III	Any Analyser (Automation / Semi Automation)	151	113.39	4.43	5.02	114 mg/dL	0.12	0.82	

DI Range	Interpretation
to +1.00	Excellent.
to ±2.00	Good.
to ±2.99	Accept with caution. Warning Signal.
	Unacceptable performance. Action Signal.

Quality Methode	2023-09-20 10:00:00	2023-09-20 10:00:00
Quality Password	123456	123456