

LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Konapet  
 Lab No: 12522  
 Month: July  
 Year: 2023  
 Constituent: Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 18/07/2023

Date of Report Published : 07/08/2023

Analyte	Method	Instrument	1	2	3	4	5	6	7	8
GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	110 mg/dL	0.11	0.28	
CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	5.8 mg/dL	-0.09	0.02	
CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation )	148	112.78	4.33	4.88	119 mg/dL	1.27	0.80	

SDI Range	Interpretation
in -1.00 to +1.00	Excellent.
in +1.01 to ±2.00	Good.
in ±2.01 to ±2.99	Accept with caution. Warning Signal.
and ±3.0	Unacceptable performance. Action Signal.


*(Handwritten Signature)*  
 11/08/2023  
 07/08/2023  
 322 503