

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Konapet
Lab No 12522
Month June
Year 2023
Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered: 19/06/2023

Date of Report Published: 04/07/2023

1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	285 mg/dL	-0.11
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.6 mg/dL	0.18
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	149	114.61	4.05	4.64	112 mg/dL	-0.58

SDI Range	Interpretation
within -1.00 to +1.00	Excellent
within +1.01 to +2.00	Good
within +2.01 to +2.99	Accept with caution - Warning Signal
beyond +3.0	Unacceptable performance - Action Signal

As