

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2023



PC-1024

PHC, Konapet	Lab No	12522
Chemistry III	Date of Result Entered :	20/05/2023
Unphosphilized human serum based	Date of Report Published :	05/06/2023

Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	SI
				CV	SD			
GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	186 mg/dL	0.55	0.44
Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.7 mg/dL	-0.45	0.01
CHOD-PAP III	Any Analyser (Automation / Semi Automation)	154	114.84	5.08	5.83	116 mg/dL	0.20	0.94

Range	Interpretation
	Excellent.
	Good.
	Accept with caution. Warning Signal.
	Unacceptable performance. Action Signal.

Coordinator Contact Details:
 clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****

[Signature]
 மருத்துவ சித்திரம்.
 மருத்துவ கல்வி நிறுவனம்
 வெள்ளூர் - 02 - 503