

Govt. PHC, Rangium

[Details About Robust Analysis](#)

12524

[Detail About Monthly Summary](#)

April

[Detail about SDI](#)

2023

Group **Chemistry III**

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 22/04/2023

Date of Report Published : 06/05/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value
					CV	SD	
URE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	127 mg/dL
URICINE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.2 mg/dL
CHOLESTEROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	143	111.22	4.31	4.79	114 mg/dL

SDI Range	Interpretation
00 to +1.00	Excellent.
01 to ±2.00	Good.
01 to ±2.99	Accept with caution. Warning Signal.
≥3.0	Unacceptable performance. Action Signal.

	Modify Methods	Submit Lab Results	View Lab Monthly Summary
	Change Password	Home	Sign-out