

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Rengam
Lab No: 12824
Month: May
Year: 2023
Constituent Group: Chemistry II

[Detail About Request Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 26/05/2023

Date of Report Published : 05/06/2023

Sl. No	Analyte	Method / Principle Name	Analyser Name	No of Participants	CV	Participants		True Value	SDI	U
						CY	SD			
1	GLUCOSE II	GOD-POD II	Robosik (Semi automatic / Fully Automatic)	1824	180.88	3.15	7.25	180 mg/dL	0.50	0.44
2	CREATININE II	Jaffes kinetic - Alkaline Phosphate II	Robosik (Semi automatic / Fully Automatic)	1824	1.78	7.48	0.12	1.7 mg/dL	-0.48	0.01
3	CHOLESTEROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	154	114.84	3.08	5.82	110 mg/dL	0.52	0.94

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to +2.00	Good.
Within +2.01 to +2.99	Accept with caution. Warning Signal.
Beyond +3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Monthly Summary	Change Password	Home	Logout