

VIEW LAB MONTHLY SUMMARY

Lab Name: **Govt. PNC, Bangalore**
 Lab No: **12524**
 Month: **July**
 Year: **2023**
 Consultant Group: **Chemistry III**

[Detail About Request Analyte](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 26/07/2023

Date of Report Published : 07/08/2023

Sl. No	Analyte	Method / Principle Name	Analyser Name	No. of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOO-POD III	Robark (Semi automatic / Fully Automatic)	1976	129.20	3.33	0.07	111 mg/dL	0.27	0.29
2	CREATININE III	Jaffes Kinetic - Alkaline Phosfat III	Robark (Semi automatic / Fully Automatic)	1830	1.88	0.00	0.33	3.8 mg/dL	-0.29	0.02
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	148	112.70	4.33	4.89	118 mg/dL	0.00	0.80

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to +2.00	Good.
Within +2.01 to +2.99	Accept with caution. Warning Signal.
Beyond +3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Signout