

VIEW LAB MONTHLY SUMMARY

Lab Name: **Dev. PwC, Bengaluru**

[Click About Request Analysis](#)

Lab No: **12204**

[Click About Monthly Summary](#)

Month: **August**

[Click about SDI](#)

Year: **2023**

Constituent Group: **Chemistry II**

[Click on the analyte to view Graphical Data](#)

[All Analyser Results](#)

[Print](#)

Date of Result Entered: **21/08/2023**

Date of Report Published: **01/09/2023**

Sl. No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Final Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Robotic (5 yrs automatic / Fully Automatic)	1841	175.00	3.13	5.00	152 mg/dL	-1.91	0.42
2	CREATININE II	Jaffes Reaction - Alkaline Phos II	Robotic (5 yrs automatic / Fully Automatic)	1852	1.78	8.77	0.12	1.5 mg/dL	1.18	0.01
3	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / 5 yrs Automation)	148	113.00	4.26	4.97	208 mg/dL	-1.20	0.81

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to +2.00	Good.
Within +2.01 to +2.99	Accept with caution. Warning Signal.
Beyond +3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Forgot Summary	Change Password	Home	Sign-out