

[VIEW LAB MONTHLY SUMMARY](#)

Lab Name: Govt. P.H.C. Ranglun

[Details About Individual Analytes](#)

Lab No: 12524

[Detail About Monthly Summary](#)

Month: September

[Detail about SDI](#)

Year: 2023

Consultant Group: Chemistry II

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 21/09/2023

Date of Report Published : 03/10/2023

Sl. No.	Analyte	Method / Principle Name	Analyser Name	No. of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD II	Robosk ( Semi automatic / Fully Automatic )	1870	310.00	4.82	14.98	318 mg/dL	0.46	0.89
2	CREATININE III	Jaffe's Kinetic - Alkaline Phosphate II	Robosk ( Semi automatic / Fully Automatic )	1832	1.25	8.14	0.10	1.3 mg/dL	0.10	0.00
3	CHOLESTROL III	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	151	113.25	4.43	5.02	109 mg/dL	-0.87	0.82

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to +2.00	Good.
Within +2.01 to +2.99	Accept with caution. Warning Signal.
Beyond +3.0	Unacceptable performance. Action Signal.

<a href="#">Links</a>			
<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">View Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Logout</a>