

**CALIBRATION / TESTING REPORT**

<b>CUSTOMER DETAILS :</b> GOVT. PRIMARY HEALTH CENTRE KULAMANGALAM, SIVAGANGAL		CERTIFICATE NO	TICS/2023/MM/SG/KUM/02
		EQUIPMENT NAME	CENTRIFUGE
		MANUFACTURE	LABOTECH
		MODEL	-
CALIBRATION DATE	03/10/2023	ASSET NO	KUMLAB-002
CALIBRATION DUE DATE	02/10/2024	INSTALLED DEPARTMENT	LAB

- |                                    |                              |  |
|------------------------------------|------------------------------|--|
| 1. General maintenance required    | <input type="checkbox"/> Yes | <input checked="" type="checkbox"/> No |
| 2. Any Physical damage observed    | <input type="checkbox"/> Yes | <input checked="" type="checkbox"/> No |
| 3. Spare Replacement (if required) | <input type="checkbox"/> Yes | <input checked="" type="checkbox"/> No |

TEST EQUIPMENT	MANUFACTURE	MODEL	SERIAL NUMBER
DIGITAL TACHOMETER	KUSAM MECO	KM2234 BL	S1049875
ELECTRICAL SAFETY ANALYZER	DATREND	4 P20 mini	100018040011

Temperature 25°C  
Humidity RH 55%

**TEST REPORT**

PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
ROTATION PER MINUTE	RPM	Speed 2	2205	± 5	PASS
		Speed 4	2856		

\*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.

ACCEPTABLE FOR USE     
  CORRECTIVE MAINTENANCE REQUIRED     
  REMOVED FROM USE

Calibrated By :




Authorized Sign With Seal

