

**CALIBRATION / TESTING REPORT**

<b>CUSTOMER DETAILS :</b> GOVT.PRIMARY HEALTH CENTRE KULAMANGALAM, SIVAGANGAL.		<b>CERTIFICATE NO</b> TICS/2023/MM/SG/KUM/01
		<b>EQUIPMENT NAME</b> MICROSCOPE
		<b>MANUFACTURE</b> MODEL
<b>CALIBRATION DATE</b>	03/10/2023	<b>ASSET NO</b> LUM LAB 001
<b>CALIBRATION DUE DATE</b>	02/10/2024	<b>INSTALLED</b> DEPARTMENT LAB

- General maintenance required  Yes  No
- Any Physical damage observed  Yes  No
- Spare Replacement (if required)  Yes  No

TEST EQUIPMENT	MANUFACTURE	MODEL	SERIAL NUMBER
ELECTRICAL SAFTY ANALYZER	DATREND	V Pad-mini	VPM 18040011

PARAMETER	UNITS	VALUE	LIMITS	STATUS
Main Voltage L1-L2	V	231.2	240±10	PASS
Main Voltage L1-G(PE)	V	232.1	240±10	PASS
Main Voltage L2-G(PE)	V	0.9	<5	PASS
Load Current	A	0.294	As per unit	PASS
Protective Earth	Ω	0	<0.3	PASS
Point-to-Point Resistance	Ω	0	<0.5	PASS
Equipment Leakage	μA	0.10	Class I - < 500 Class II - < 100	PASS
Applied Part Leakage	μA	0	Type BF - < 5mA Type CF- < 50μA*	PASS

Note : Test Standard: IEC 62353  
Defibrillator paddles 100 μA, per IEC 60601-2-4

\*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.

ACCEPTABLE FOR USE  CORRECTIVE MAINTENANCE REQUIRED  REMOVED FROM USE

Calibrated By :  
*Pr*

Authorized Sign With Seal