

CALIBRATION / TESTING REPORT

CUSTOMER DETAILS : GOVT.PRIMARY HEALTH CENTRE KULAMANGALAM, SIVAGANGAL.		CERTIFICATE NO	TICS/2023/MM/SG/KUM/04
		EQUIPMENT NAME	MICROPIPETTE
		MANUFACTURE	THERMO SCIENTIFIC
		MODEL	10-100 µL
CALIBRATION DATE	03/10/2023	ASSET NO	RW01480
CALIBRATION DUE DATE	02/10/2024	INSTALLED DEPARTMENT	LAB

1. Have more dust in machine outside Yes No
2. Any Physical damaged in machine Yes No
3. Machine inside Cleaning (if required) Yes No

TEST EQUIPMENT	ID. NO	CERTIFICATE NUMBER	VALIDITY
ELECTRONIC SEMI MICRO BALANCE	NCS/WB01	TSC/23-24/5424-1	27-JUNE-2024

TEST REPORT

Temperature 25°C
Humidity RH 55%

PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	50	50.12	± 1	PASS
		100	999.86	± 1	PASS

*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.

ACCEPTABLE FOR USE
 CORRECTIVE MAINTENANCE REQUIRED
 REMOVED FROM USE
 Calibrated By : _____ Authorized Sign With Seal

Per



[Signature]