

CALIBRATION / TESTING REPORT

| | | | |
|---|------------|----------------------|---------------------------------|
| CUSTOMER DETAILS : GOVT.PRIMARY HEALTH CENTRE KULAMANGALAM, SIVAGANGAI. | | CERTIFICATE NO | TICS/2023/MM/SG/KUM/05 |
| | | EQUIPMENT NAME | MICROPIPETTE |
| | | MANUFACTURE MODEL | THERMO SCIENTIFIC 100-1000µL |
| | | SERIAL NO | RW02956 |
| CALIBRATION DATE | 03/10/2023 | INSTALLED DEPARTMENT | LAB |
| CALIBRATION DUE DATE | 02/10/2024 | | |

- 1. Have more dust in machine outside Yes No
- 2. Any Physical damaged in machine Yes No
- 3. Machine inside Cleaning (if required) Yes No

| TEST EQUIPMENT | ID. NO | CERTIFICATE NUMBER | VALIDITY |
|-------------------------------|----------|--------------------|--------------|
| ELECTRONIC SEMI MICRO BALANCE | NCS/WB01 | TSC/23-24/5424-1 | 27-JUNE-2024 |

TEST REPORT

Temperature 25°C
Humidity RH 55%

| PARAMETERS | UNITS | SET VALUE | MEASURED VALUES | TOLERANCE | STATUS |
|------------|-------|-----------|-----------------|-----------|--------|
| VOLUME | µL | 100 | 100.08 | ± 1 | PASS |
| | | 1000 | 999.98 | ± 1 | PASS |

*The above mentioned instrument has been tested using standards manufactures recommended protocols. using equipment having traceability to National/International standards.

ACCEPTABLE FOR USE
 CORRECTIVE MAINTENANCE REQUIRED
 REMOVED FROM USE
 Authorized Sign With Seal

Calibrated By :
Per



[Signature]