

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Kalamangalam

Lab No: 10802

Month: September

Year: 2023

Constituent Group: Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyze to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 18/09/2023

Date of Report Published : 05/10/2023

Sl.No	Analyze	Method / Principle Name	Analyser Name	No of Participants	CV	Participants		Your Value	SDI	I
						CV	SD			
1	GLUCOSE III	GGD-POD III	Roburisk Semi automatic / Fully Automatic	1870	318.88	4.82	14.98	325 mg/dL	-0.28	0.68
2	CREATININE III	Jaffee Kinetic - Alkaline Phorate II	Roburisk Semi automatic / Fully Automatic	1852	1.28	8.14	0.18	1.8 mg/dL	2.95	0.08
3	CHOLESTEROL III	CHOD-PAP III	Roburisk Semi automatic / Fully Automatic	1782	113.82	3.88	4.82	98 mg/dL	-3.62	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Weekly Summary	Change Password	Home	Sign out