

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Kalamangalam
Lab No 10802
Month August
Year 2023
Constituent Group Chemistry III

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

[Click on the analysis to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 12/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	CV	Participants		Your Value	SDI	I
						CV	SD			
1	GLUCOSE III	GDH-POD II	Robosk Semi automatic / Fully Automatic	1841	175.58	5.15	9.80	170 mg/dL	-0.62	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Phosphate II	Robosk Semi automatic / Fully Automatic	1852	1.7%	6.77	0.12	1.7 mg/dL	-0.58	0.01
3	CHOLESTROL III	CHOD-PAP II	Robosk Semi automatic / Fully Automatic	1764	113.89	4.87	4.64	111 mg/dL	0.64	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Verify Summary	Change Password	Home	Logout