

## VIEW LAB MONTHLY ALL RESULT SUMMARY

Lab Name **Govt. PHC, Mathiyallur**Lab No **12505**[Back to Search Page](#)Month **September**Year **2023**[Print](#)Constituent Group **Chemistry III**Date of Result Entered : **20/09/2023**Date of Report Published : **05/10/2023**

Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	36	312.41	4.03	12.59			4.20
			Transasia / Erba	18	307.20	6.79	20.85			9.83
			Robonik ( Semi automatic / Fully Automatic )	1870	310.86	4.82	14.98	294 mg/dL	-1.13	0.69
			Any Analyser (Automation / Semi Automation )	68	312.35	4.82	15.05			3.65
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	187	35.19	12.73	4.48	34 mg/dL	-0.27	0.66
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	38	1.30	7.23	0.09			0.03
			Transasia / Erba	21	1.26	9.52	0.12			0.05
			Robonik ( Semi automatic / Fully Automatic )	1852	1.29	8.14	0.10	1.2 mg/dL	-0.86	0.00
			Any Analyser (Automation / Semi Automation )	69	1.31	8.62	0.11			0.03
4	CHOLESTROL III	CHOD-PAP III	Biosystems	35	113.55	3.41	3.87			1.31
			Transasia / Erba	19	113.64	2.93	3.33			1.53
			Robonik ( Semi automatic / Fully Automatic )	1783	113.62	3.80	4.32	110 mg/dL	-0.84	0.20
			Any Analyser (Automation / Semi Automation )	151	113.39	4.43	5.02			0.82

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

