

## VIEW LAB MONTHLY ALL RESULT SUMMARY

Lab Name **Govt. PHC, Mathiyallur**Lab No **12505**[Back to Search Page](#)Month **May**Year **2023**[Print](#)Constituent Group **Chemistry III**Date of Result Entered : **20/05/2023**Date of Report Published : **05/06/2023**

Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	33	179.81	5.48	9.85			3.43
			Transasia / Erba	19	177.53	5.62	9.98			4.58
			Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	175 mg/dL	-0.63	0.44
			Any Analyser (Automation / Semi Automation )	69	181.20	5.81	10.52			2.53
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	163	27.91	13.39	3.74	27 mg/dL	-0.24	0.59
			Any Analyser (Automation / Semi Automation )	11	25.79	11.91	3.07			1.85
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	38	1.78	8.91	0.16			0.05
			Transasia / Erba	22	1.85	7.73	0.14			0.06
			Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.7 mg/dL	-0.45	0.01
			Any Analyser (Automation / Semi Automation )	68	1.75	7.05	0.12			0.03
4	CHOLESTROL III	CHOD-PAP III	Biosystems	36	116.50	6.42	7.47			2.49
			Transasia / Erba	20	115.73	3.06	3.54			1.58
			Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	99 mg/dL	-2.68	0.27
			Any Analyser (Automation / Semi Automation )	154	114.84	5.08	5.83			0.94

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

## Links

[Modify Address](#)[Modify Methods](#)[Submit Lab Results](#)[View Lab Monthly Summary](#)[Yearly Summary](#)[Change Password](#)[Home](#)[Sign-out](#)

