



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - AUGUST 2023



PG-10/24

Lab Name: Govt. PHC, Kandachipuram
 Constituent Group: Chemistry III
 PT Item: Lyophilized human serum based

Lab No: 13110
 Date of Result Entered: 19/08/2023
 Date of Report Published: 01/09/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	174 mg/dL	-0.18	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.7 mg/dL	-0.50	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	115 mg/dL	0.26	0.22
Interpretation										
SDI Range				Excellent.						
Within -1.00 to +1.00				Good.						
Within ±1.01 to ±2.00				Accept with caution. Warning Signal.						
Within ±2.01 to ±2.99				Unacceptable performance. Action Signal.						
Beyond ±3.0										

LAB ADDRESS :
 Govt. PHC, Kandachipuram
 Kandachipuram
 Kandachipuram
 TAMILNADU605701

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christodoss
MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENTRE
 KANDACHIPURAM
 WILLUPURAM DISTRICT.

Dr. Pamela Christodoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential