



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - JUNE 2023



PC 1021

Lab Name: Govt. PHC, Kandatchipuram
 Constituent Group: Chemistry III
 PT Item: Lyophilized human serum based

Lab No: 13110
 Date of Result Entered: 20/06/2023
 Date of Report Published: 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	292 mg/dL	0.36	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.7 mg/dL	0.76	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	115 mg/dL	0.25	0.24
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt. PHC, Kandatchipuram
 kandatchipuram
 Kandatchipuram
 TAMILNADU605701

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENTRE
 KANDATCHIPURAM
 VILUPURAM DISTRICT.

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential