

HOPE ULTRA SYSTEMS

No: 2/79, MAIN ROAD, MAMSAPURAM (PO),
SIVAKASI, VIRUDHUNAGAR-626124.
E-mail : hopeultrasystems@gmail.com

GSTIN. 33DLMPK3057D1Z9
Mobile: +919487402761, +919715223192.

CALIBRATION CERTIFICATE

Requested By
GPHC, M.REDDIAPATTI.

Scope:		Asset no	: PHC/MRP-LAB/002
Equipment Name	: Micropipette	Calibration Done at	: customer place
Make	: Thermo scientific(Finnpipette F3)	Calibration Date	: 05/09/2023
Model	: 100-1000 μ l	Next call due on	: 04/09/2024
Sl no	: RW02684		

TEST RESULTS Volume μ l

Sl. No.	DUC reading(μ l)	Standard Reading (μ l)		Correction (μ l)
		Trail 1	Trail 2	
01	100	100.0	100.0	+0
02	200	200.0	200.0	+0
03	500	500.0	500.0	+0
04	800	800.0	800.0	+0
05	1000	1000.0	1000.0	+0

Error Acceptance : Not Applicable (user defined)

Equipment Status : Working in Good condition

Materials used for testing and calibration:

- Electronic weighing scale(EWC/SF400C)

Note:

- Calibration points were selected as per customer request.
- Each reading was taken by repeatability.
- Results reported are valid at the time of and under the stated conditions of measurements.

Hope Ultra Systems

Authorized signatory

