



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name: Govt. PHC, Kavanur

Lab No: 12552

Constituent Group: Chemistry III

Date of Result Entered: 19/06/2023

PT Item: Lyophilized human serum based

Date of Report Published: 04/07/2023

S.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	291.2 mg/dL	0.31	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.65 mg/dL	0.47	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	116.8 mg/dL	0.61	0.24
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. PHC, Kavanur
Kavanur
TAMILNADU

Coordinator Contact Details:
Email: cingqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christodoss

Dr. Pamela Christodoss
CMC EQAS Coordinator
Christian Medical College, Vellore

R. Kavanur

Assistant Surgeon
Govt. Primary Health Centre,
R. Kavanur,
Ramanathapuram Dist.

Homogeneity and Stability of the sample is passed.
Data In CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****