

HOPE ULTRA SYSTEMS

No: 2/79, MAIN ROAD, MAMSA PURAM (PO),
SIVAKASI, VIRUDHUNAGAR-626124.
E-mail : hopeultrasystems@gmail.com

GSTIN. 33DLMPK3057D1Z9
Mobile: +919487402761, +919715223192.

CALIBRATION CERTIFICATE

Requested By

GPHC, M. REDDIAPATTI.

Scope:

Equipment Name : Micropipette	Asset no : PHC/MRP-LAB/003
Make : Thermo scientific(Finnpipette F3)	Calibration Done at : customer place
Model : 10-100 μ l	Calibration Date : 05/09/2023
Sl no : PW02822	Next call due on : 04/09/2024

TEST RESULTS Volume μ l

Sl. No.	DUC reading(μ l)	Standard Reading (μ l)		Correction (μ l)
		Trail 1	Trail 2	
01	10	10.0	10.0	+0
02	20	20.0	20.0	+0
03	50	50.0	50.0	+0
04	80	80.0	80.0	+0
05	100	100.0	100.0	+0

Error Acceptance : Not Applicable (user defined)

Equipment Status : Working in Good condition

Materials used for testing and calibration:

- Electronic weighing scale(EWC/SF400C)

Note:

- Calibration points were selected as per customer request.
- Each reading was taken by repeatability.
- Results reported are valid at the time of and under the stated conditions of measurements.

Hope Ultra Systems

