



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Lab Name: Govt. PHC, Kavanur

Constituent Group: Chemistry III

PT item: Lyophilized human serum based

Lab No: 17552

Date of Result Entered: 15/04/2023

Date of Report Published: 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	119.1 mg/dl	-0.05	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.4 mg/dl	1.37	0.61
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	110.9 mg/dl	-0.12	0.28
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. PHC, Kavanur

Kavanur

TAMILNADU

Coordinator Contact Details:

Email: clinqc@cmcwellore.ac.in

Contact Number: 0416-2283102

Pavala Christadevi

Dr. Pamela Christudoss

CMC EQAS Coordinator

Christian Medical College, Vellore

Assistant Surgeon

Govt. Primary Health Centre

R. Kavanur,

Ramanathapuram Dist

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****