

## CENTRIFUGE Calibration / Verification Record Sheet

Location: ICTC- DH- HINDUPUR

Range : Up to 3500 RPM

Equipment ID: APSACS/ICTC/DH/HDP/CEN/01

SL. No:02

Calibration Procedure Reference: In house verification of centrifuge using standard calibrated by Tachometer

Date of Calibration	11-10-2023	Due Date of Calibration	10-10-2024	
Certified Centrifuge ID: GMC-ATP/HIVTL/SRL/CEF/01	Tachometer ID: GMC-ATP/HIVTL/SRL/TM/01	Calibration date: 19-05-2023	Due date of Calibration: 18-05-2024	
Verification Done By: P. ANURADHA, Technical Officer, SRL, GMC, Ananthapuramu				
No.	STANDARD READING (RPM)	UUC READING (RPM)	ERROR (RPM)	
1.	499.1	500	0.9	
2.	999.5	1000	0.5	
3.	1998.7	2000	1.3	
4.	2997.8	3000	2.2	
5.	3496.4	3500	3.6	

Accept/ Reject: Accepted since average difference is within +/-1000RPM

Action Taken: NA

Remarks: NA

P. Anuradha  
SRL-1.0  
DEPT. of Microbiology  
GMC - Ananthapuramu.