

Certificate No: PHCO/002



**BIOMEDICAL  
TECHNIQUES**  
Calibration experts

### CALIBRATION CERTIFICATE

Customer Details		Details of DUC	
GOVT PRIMARY HEALTH CENTRE, OLAVILAI, KANNIYAKUMARI, TAMIL NADU- 629702	Description	BIOCHEMISTRY ANALYZER	
	Manufactured by	ROBONIK	
	Model	PRIETEST	
	Serial No.	AT1061010RBK	
	Identification No.	NIL	
Location		LABORATORY	

Environmental Conditions		Calibration Details	
Operating Temperature	28°C	Date of Receipt	07/03/2023
Humidity	55 %	Date of Calibration	07/03/2023
		Date of Issue	10/03/2023
		Due Date of calibration	06/03/2024
		Calibrated At	Site

#### Details of Calibration Standards

Name	Manufacturer	Serial No.	Valid up to	Traceability
Electrical safety analyser	Fluke / ESA 615	2244039	25-SEP-2023	Accuver calibration laboratory CO LLC, Dubai

#### Electrical Safety Test

Parameter	Measured Value	Tolerance	Result
Ground wire Resistance	0.156 Ω	<0.3 Ω	Pass
insulation Resistance	OL	> 2 M Ω	Pass
Chassis Leakage Current *NC	0.6 μA	<100 μA	Pass
Chassis Leakage Current *SFC	0.4 μA	<500 μA	Pass
Patient Leakage Current - B	0.8 μA	<100μA	Pass

Ph: +91-487-24 25 821  
+91-98 95 02 58 21  
+91- 90 48 44 51 11

biomedicaltechniques@gmail.com

Registered Office : II/446, S N Park, Poothole PO,  
Thrissur, Kerala, India - 680 004

Corporate Office - XXV/277, Unity Building  
Mannadiar Lane, Thrissur, Kerala - 680001

QUALITY MANAGEMENT ISO 9001 : 2008

www.bmtqa.com