



Certificate No : PHCT/MP/002

## CALIBRATION CERTIFICATE

Customer Details		Details of DUC	
GOVT PRIMARY HEALTH CENTRE, OLAVILAI TAMILNADU-629174	Description	MICROPIPETTE	
	Manufactured by	THERMO SCIENTIFIC	
	Model	FINN PIPETTE F3	
	Serial No.	RW01400	
	Identification No.	-	
	Range	10-100 $\mu$ L	
	Location	LABORATORY	

Environmental Conditions		Calibration Details	
Operating Temperature	23°C	Date of Receipt	17/06/2023
		Date of Calibration	21/06/2023
		Date of Issue	24/06/2023
Humidity	55 %	Rec. Due Date	23/12/2023
		Calibrated At	Site
		Cond. On Receipt	Satisfactory

### Details of Calibration Standards

Name	Make/Model SI No	Valid up to	Traceability
Electronic Weighing Balance	AND,GR-202/ 14251938	30/09/2023	Equations Calibration Services,Chennai

### MECHANICAL CALIBRATION

(Volume)

Parameter	Set Value( $\mu$ l)	Measured Value( $\mu$ l)	systematic Error( $\pm\mu$ l)
Volume	10	10.1	0.1
	50	49.8	0.2
	100	100.2	0.2