

Certificate No : PICT/MP/003

CALIBRATION CERTIFICATE

Customer Details	Details of DUC	
GOVT PRIMARY HEALTH CENTRE, OLAVILAI TAMILNADU-629174	Description	MICROPIPETTE
	Manufactured by	THERMO SCIENTIFIC
	Model	FINN PIPETTE F3
	Serial No.	RW02571
	Identification No.	-
	Range	100-1000 μ L
	Location	LABORATORY

Environmental Conditions		Calibration Details	
Operating Temperature	23 $^{\circ}$ C	Date of Receipt	17/06/2023
		Date of Calibration	21/06/2023
		Date of Issue	24/06/2023
Humidity	55 %	Rec. Due Date	23/12/2023
		Calibrated At	Site
		Cond. On Receipt	Satisfactory

Details of Calibration Standards

Name	Make/Model SI No	Valid up to	Traceability
Electronic Weighing Balance	AND,GR-202/ 14251938	30/09/2023	Equations Calibration Services,Chennai

MECHANICAL CALIBRATION

(Volume)

Parameter	Set Value(μ l)	Measured Value(μ l)	systematic Error($\pm\mu$ l)
Volume	100	100.3	0.3
	500	499.8	0.2
	1000	1000.2	0.2