

## Calibration Report

<b>Customer Address:</b> The Medical officer, Govt, Primary Health Centre, Bagepalli, Krishnagiri District.			<b>Date</b>		09.07.2022	
			<b>Machine SI No</b>			
			<b>MasterSample Calibration date</b>		21.02.2022	
S.No	Product	Spec	Range	Warranty	Problem Observed	Solution Provided
1	Autoclave	150 Kg/cm <sup>2</sup> per 45 to 60 mins	Constant	Yes	Nil	Nil
2	ECG Machine	UPL DLL	Constant	Yes	Nil	Nil
3	Pluse Oxy Meter	80 To 100	85 86	Yes	Nil	Nil
4	Baby Wammer	34.5 to 36.00	34.80 35.00	Yes	Nil	Nil
5	Deep Freezer	-2.00 to -25.00	-19.50	Yes	Nil	Nil
6	ILR Machine	+2.0 To +8.0	+3.9 +4.0	Yes	Nil	Nil
7.	O2 Concentrator	Flow 2.5L/min	2.5L/m	Yes	Nil	Nil
8.	Micro Pipette	Suction 500 µl/ 1000 µl	500 µl/ 1000 µl	Yes	Nil	Nil
9.	Semi auto analyzer	Lamp Amtu Above 3.0	4.8 5.2	Yes	Nil	Nil
10.	Centrifuge	2800 RPM	T1 -2400 T2- 2800	Yes	Nil	Nil
11	Weighing machine	+0.5 -0.5	10kg T1 50kg T2	Yes	Nil	Nil

1. Annual maintenance contract for **Bagepalli** PHC, 2. AMC Period 1 Year, from 09.07.22. to 08.07.23.

3. Company shall provide service from 17.02.22 to 16.02.23. Feel free to contact our service engineer

For ALLWIN MEDICAL SYSTEMS

