



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name

Govt. PHC, Muttal

Lab No

12621

Constituent Group

Chemistry III

Date of Result Entered :

20/09/2023

PT Item

Lyophilized human serum basod

Date of Report Published :

05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	306.2 mg/dL	-0.31	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.3 mg/dL	0.10	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	107.2 mg/dL	-1.49	0.20
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. PHC, Muttal
MUTTAL,SALEM
Muttal
TAMILNADU636121

Coordinator Contact Details:

Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

[Signature]
MEDICAL OFFICER
Govt. Primary Health Centre
Muttal-636 112

[Signature]
Dr. Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore