



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JULY 2023

Lab Name: Govt. PHC, Poomparai
Constituent Group: Chemistry III
PT Item: Lyophilized human serum based

Lab No:
Date of Result Entered:
Date of Report Published:

SLNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	CV	Participants		Your Value
						CV	SD	
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	99.8 mg/dL
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	6.83	6.08	0.35	5.4 mg/dL
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	115 mg/dL

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS:
Govt. PHC, Poomparai
THE MEDICAL OFFICER, KODAIKANAL
Poomparai
TAMILNADU

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0415-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential.
CMC EQAS does not sub contract any components
***** End of Report *****