



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JUNE 2023



Lab Name: Govt. PHC, Poomparai  
Constituent Group: Chemistry III  
PT Item: Lyophilized human serum based

Lab No: 12188  
Date of Result Entered: 26/06/2023  
Date of Report Published: 04/07/2023

S.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Mean Value	SD	CV
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1023	200.03	5.17	14.82	230.3 mg/dL	5.23	6.88
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1048	2.57	6.64	0.17	2.5 mg/dL	1.82	5.87
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1004	113.73	4.48	6.07	116.4 mg/dL	6.80	5.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

#### LAB ADDRESS :

Govt. PHC, Poomparai  
THE MEDICAL OFFICER, KODAIKANAL  
Poomparai  
TAMILNADU

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudore*  
Dr. Pamela Christudore  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*