


CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS : GOVT.PRIMARY HEALTH CENTRE MONNI KARMANKUDI, SIVAGANGAI.		CERTIFICATE NO		TICS/2023/SG/MKK/002	
		EQUIPMENT NAME		CENTRIFUGE	
		MANUFACTURE		LAB TECH	
		MODEL		--	
CALIBRATION DATE	22/09/2023	ASSET NO		MKK PH 002	
CALIBRATION DUE DATE	21/09/2024	INSTALLED DEPARTMENT		LAB	
1. General maintenance required <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damage observed <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Spare Replacement (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No					
TEST EQUIPMENT		MANUFACTURE		MODEL	
DIGITAL TACHOMETER		KUSAM MECO		KM2234 BL	
ELECTRICAL SAFTY ANALYZER		DATREND		V Pad-mini	
				SERIAL NUMBER	
				S1049875	
				VPM 18040011	
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
ROTATION PER MINUTE	RPM	SPEED 2	1551	± 5	PASS
		SPEED 4	2096		
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By : 					
Authorized Sign With Seal 