



CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS : GOVT.PRIMARY HEALTH CENTRE MONNI KARMANKUDI, SIVAGANGAI.		CERTIFICATE NO		TICS/2023/SG/MKK/04	
		EQUIPMENT NAME		MICROPIPETTE	
		MANUFACTURE MODEL		THERMO SCIENTIFIC 10-100µL	
		SERIAL NO		MKK PH 004	
CALIBRATION DATE	22/09/2023	INSTALLED DEPARTMENT		LAB	
CALIBRATION DUE DATE	21/09/2024				
<p>1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p>					
TEST EQUIPMENT	ID. NO	CERTIFICATE NUMBER		VALIDITY	
ELECTRONIC SEMI MICRO BALANCE	NCS/WB01	TSC/23-24/5424-1		27-JUNE-2024	
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	10	10.03	± 1	PASS
		100	99.99	± 1	PASS
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By :		Authorized Sign With Seal			