




CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS : GOVT.PRIMARY HEALTH CENTRE MONNI KARMANKUDI, SIVAGANGAI.			CERTIFICATE NO	TICS/2023/SG/MKK/03	
			EQUIPMENT NAME	MICROPIPETTE	
			MANUFACTURE	THERMO SCIENTIFIC	
			MODEL	100-1000µL	
CALIBRATION DATE	22/09/2023	SERIAL NO	MKK PH 003		
CALIBRATION DUE DATE	21/09/2024	INSTALLED DEPARTMENT	LAB		
<p>1. Have more dust in machine outside      <input type="checkbox"/> Yes      <input checked="" type="checkbox"/> No</p> <p>2. Any Physical damaged in machine      <input type="checkbox"/> Yes      <input checked="" type="checkbox"/> No</p> <p>3. Machine inside Cleaning (if required)      <input type="checkbox"/> Yes      <input checked="" type="checkbox"/> No</p>					
TEST EQUIPMENT	ID. NO	CERTIFICATE NUMBER	VALIDITY		
ELECTRONIC SEMI MICRO BALANCE	NCS/WB01	TSC/23-24/5424-1	27-JUNE-2024		
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	100	100.11	± 1	PASS
		1000	1000.08	± 1	PASS
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By : 			Authorized Sign With Seal 