



VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Chinnankuppam [Details About Robust Analysis](#)

Lab No: 12106 [Detail About Monthly Summary](#)

Month: July [Detail about SDI](#)

Year: 2023

Constituent Group: Chemistry III

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 21/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation )	75	109.71	5.43	5.95	108 mg/dL	-0.29	1.37
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation )	66	5.89	4.60	0.27	5.8 mg/dL	-0.33	0.07
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation )	148	112.78	4.33	4.88	110 mg/dL	-0.57	0.80

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

Site Maintained by **Swiftware Solutions**

Best Viewed at pixels 1024 X 768

