



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name	Govt. PHC, Porperappi	Lab No	12015
Constituent Group	Chemistry III	Date of Result Entered:	20/09/2023
PT Item	Lyophilized human serum based	Date of Report Published:	05/10/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	SI
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robosk (Semi automatic / Fully Automatic)	1870	310.80	4.92	14.98	305 mg/dL	-0.39	0.09
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robosk (Semi automatic / Fully Automatic)	1802	1.29	8.14	0.10	1.2 mg/dL	-0.86	0.00
3	CHOLESTROL III	CHOD-PAP III	Robosk (Semi automatic / Fully Automatic)	1783	113.00	3.80	4.32	109 mg/dL	-1.07	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to +2.00	Good
Within +2.01 to +3.99	Accept with caution. Warning Signal.
Beyond +3.0	Unacceptable performance. Action Signal.

LAB ADDRESS:  
Govt. PHC, Porperappi  
THE MEDICAL OFFICER, SENIURAI  
Porperappi  
TAMILNADU

Coordinator Contact Details:  
Email: clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and stability of the sample is passed  
Data in CMC EQAS reports is confidential  
CMC EQAS does not subcontract any components  
\*\*\*\*\* End of Report \*\*\*\*\*